



Limited liability company

«Biolabmix»

TIN 5408278957 CAT 540801001

630090, Novosibirsk obl., Novosibirsk,

st. Injenernaya, building № 28

Tel/Fax: +7(383)363-51-91, Tel: +7(383)363-22-40

E-mail: sales@biolabmix.ru

## BioMaster UDG HS-qPCR Lo-ROX (2\*)

Cat. number MHR023-400, MHR023-2040

### Product description:

BioMaster UDG HS-qPCR Lo-ROX (2\*) kit contains ready-to-use BioMaster UDG HS-qPCR Lo-ROX (2\*) master mix and sterile water. BioMaster UDG HS-qPCR Lo-ROX (2\*) is developed for quantitative real-time PCR with fluorescent probes on PCR platforms that use ROX as a reference dye. BioMaster UDG HS-qPCR Lo-ROX (2\*) includes all the components necessary for PCR (excluding DNA template, primers and probe):

- highly processive recombinant HS-Taq DNA polymerase;
- uracil-DNA glycosylase;
- deoxynucleoside triphosphate mix;
- PCR buffer;
- Mg<sup>2+</sup> (5 mM);
- ROX dye.

The mix is optimized for consistent and efficient real-time hot-start PCR of genomic, plasmid and viral DNA samples. The mix is supplemented with additives that increase half-life and processivity of HS-Taq DNA polymerase by enhancing its stability during PCR. BioMaster UDG HS-qPCR Lo-ROX (2\*) does not contain substances affecting primer annealing temperature and characteristics of template melting.

The presence of uracil-DNA glycosylase and dUTP (proportional to TTP) provides reliable protection against reamplification of carryover PCR products between reactions (cross-contamination). DNA polymerase included in the BioMaster UDG HS-qPCR Lo-ROX (2\*) is inactive at room temperature, and its activation requires preheating of the reaction solution at 95 °C for 5 min.

The master mix is ideally suitable for PCR platforms that use ROX passive dye as a reference guide: Life Technologies (ABI) 7500, 7500 Fast, ViiA 7, QuantStudio 12K; Stratagene Mx4000, Mx3005P, Mx3000P etc. Use of the kit saves time and minimizes contamination risk due to reduced number of pipetting steps.

### Product composition

Cat. #	BioMaster UDG HS-qPCR Lo-ROX (2*)	Water	Number of reactions (25 µl)
MHR023-400	4 × 1.25 ml	4 × 1.25 ml	400
MHR023-2040	17 × 1.5 ml	3 × 1.8 ml	2040

### BioMaster UDG HS-qPCR Lo-ROX (2\*) contains:

100 mM Tris-HCl (pH 8.5 at 25 °C), 100 mM KCl, deoxynucleoside triphosphate mix (including dUTP), 10 mM MgCl<sub>2</sub>, 0.1 U/µL HS-Taq DNA polymerase, uracil-DNA glycosylase, 0.025% Tween 20, stabilizers of HS-Taq DNA polymerase, 0.9 µM ROX fluorescent dye.

**Limits of use:**

Not recommended to use for real-time PCR with intercalating dyes. **BioMaster UDG HS-qPCR SYBR Blue (2×)** or **BioMaster UDG HS-qPCR Lo-ROX SYBR (2×)** should be used for such purposes.

**Area of application:**

- Real-time hot start PCR with fluorescently labeled probes and ROX as a reference dye;
- Conventional PCR;
- High-throughput PCR;
- Multiplex PCR;
- Genotyping.

**Polymerase features**

Recombinant *HS-Taq* DNA polymerase possesses 5'-3' DNA-dependent polymerase activity and 5'-3' exonuclease activity of native *Taq* DNA polymerase from *Thermus aquaticus*. The extension rate of *Taq* DNA polymerase depends on the complexity of DNA template and is approximately 1 kbp/min. Recombinant form of the enzyme is ideal for both conventional and real-time PCR.

**Passive fluorescent ROX dye**

The mix includes passive fluorescent ROX dye, which serves as the inner reference for signal normalization of dyes comprising oligonucleotide probes when using PCR platforms with such function (Applied Biosystems). ROX allows adjustment of variations between tubes (wells) that occur due to the pipetting errors and fluorescence fluctuation. The presence of ROX does not affect the course of PCR and shift in fluorescence signal in case if the mix is used with other PCR platforms. However, it should be taken into account that the presence of ROX fluorophore restricts its use for oligonucleotide probes, as well as for other dyes that share similar spectral characteristics ( $E_m \sim 621 \text{ nm}$ ).

**Product features:**

- The mix is optimized for real-time hot-start PCR with fluorescently labeled probes;
- Allows normalization to ROX reference dye;
- Prevents re-amplification of extraneous PCR products;
- The mix contains substances that increase its storage terms (the storage of **BioMaster UDG HS-qPCR Lo-ROX (2×)** at room temperature for 7 days does not reduce PCR efficiency) and allow multiple thawing-freezing cycles.

**Benefits of use:**

- The enzyme with hot start capability enhances reaction specificity;
- Activation of HS- *Taq* DNA polymerase requires not more than 5 min heating;
- High selectivity and reaction yield;
- Reduced preparation time;
- Protection against cross-contamination;

- Possibility of data normalization;
- Standardized conditions of the same-type reactions (reduced pipetting error during mixing PCR components in a series of experiments);
- Minimized efforts.

### Amplification protocol

1. Thaw the reaction mixture and vortex thoroughly.
2. Add the following components into thin-wall PCR tubes considering the final volume of a reaction mixture equal to 25  $\mu$ L:

Component	Volume	Final concentration
<i>BioMaster UDG HS-qPCR Lo-ROX (2x)</i>	12,5	1x
Forward primer	variable	0.1 – 600 nM
Reverse primer	variable	0.1 – 600 nM
Probe	variable	0.1 – 300 nM
DNA template	variable	1 pg – 1 $\mu$ g
Sterile water	up to 50 $\mu$ L	

3. Gently vortex and remove droplets by brief centrifugation.
4. Perform PCR, using temperature conditions recommended below:

Three-step protocol:

Step	Temperature, °C	Incubation time	Number of cycles
Anti-contamination treatment	50	2 min	1
Preliminary denaturation	95	5 min	1
Denaturation	95	5 – 15 sec	
Annealing	50 – 68	5 – 15 sec	30-50
Elongation	58 – 72	30 – 60 sec	

Or:

Two-step protocol:

Step	Temperature, °C	Incubation time	Number of cycles
Anti-contamination treatment	50	2 min	1
Preliminary denaturation	95	3 – 5-min	1
Denaturation	95	5 – 15 sec	
Annealing/elongation	50 – 68	30 – 60 sec	30-50

5. PCR result is displayed as amplification curve.

**Storage conditions:** in a place protected from light at +25 ° C - 7 days; at +4 ° C - 4 months; at -20 ° C - 18 months; not more than 50 thawing-freezing cycles.

**Transportation:** Transport in thermocontainers with cooling elements; the ambient temperature increment to the room temperature during the transportation up to 10 days is allowed.